Determination of degradation constant of Li ion from $^{10}\text{B}(n, \alpha\gamma)^7\text{Li}$ reaction in various media

By G. M. Sun^{1,*}, Y. Lee¹, J. H. Moon¹ and R. Acharya²

¹ Korea Atomic Energy Research Institute, Deokjin-dong, Yuseong-gu, Daejeon 305-353, Korea

² Bhabha Atomic Research Centre, Trombay, Mumbai 400-085, India

(Received December 22, 2009; accepted in revised form November 11, 2010)

*Doppler broadening effect / Degradation constant /
 Li ion / $^{10}\text{B}(n, \alpha\gamma)^7\text{Li}$ reaction /
 Prompt gamma activation analysis / HYPERGAM*

Summary. A Li ion from a $^{10}\text{B}(n, \alpha\gamma)^7\text{Li}$ reaction induced by a thermal neutron undergoes a large degradation due to its high stopping power in a medium. This phenomenon influences the peak line shape of the 477.6 keV gamma line emitted from an excited ^7Li nucleus recoiled by an energetic alpha particle and the peak line shape in a spectrum varies according to the slowing-down media. A numerical study for the analysis of the line shape has been conducted by performing a parametric search and fitting the measured spectrum in a least squares approach by Choi *et al.* and implemented in a recent version of HYPERGAM code. In this study, the gamma-ray spectra were measured for various media like elemental boron, boric acid, borax, boron carbide, boron steel and so on at a SNU-KAERI prompt gamma activation analysis facility of the HANARO research reactor in Korea Atomic Energy Research Institute. The degradation constant was determined for various slowing-down media.

1. Introduction

In a gamma-ray spectrum, the spectral line shape depends on the broadening mechanisms of the initial and final states, and includes various sources like a natural line width, detector response and so on. And gamma rays emitted from a recoiled nucleus can undergo a Doppler effect, which causes line width broadening. Especially for the case of prompt gamma rays from the $^{10}\text{B}(n, \alpha)^7\text{Li}^*$ reaction, broadening of a 477.6 keV line combined with a recoil of an excited Li nucleus is the characteristic of prompt gamma activation analysis (PGAA), which leads to a complex gamma-ray spectrum which is difficult to analyze. Hauser *et al.* [1] and Sakai *et al.* [2] have represented a line shape of the Doppler-broadened boron peak (DBP) by using a double numerical integration method. Kubo *et al.* [3] and Magara *et al.* [4] have developed more simplified fitting methods for the DBP and its interfered Gaussian peaks. Sun *et al.* [5] shows that a numerical fitting method of the convolution integral of the detector response function and the Doppler-broadened emission spectrum is successful in the low stopping-power

media, where the velocity of the Li ion can be treated to be constant up to the full stopping [5].

The mean life time (t) of the excited state of the ^7Li nucleus is 0.105 ps and its decay constant is given to be 9.49 ps^{-1} [6]. The ^7Li ion with an initial kinetic energy of 840 keV can fly to about 1.8 mm and 1.0 mm away from the neutron capture position, in elemental boron and tungsten, respectively. The flight time to the full stopping is comparable to or just several times larger than the decay time (t), hence most of the ^7Li ions undergo an abrupt drop of the velocity during the flight and some nuclei emit gamma rays after full stopping. The variation in velocity of the ^7Li ion influences the line shape and can be considered in terms of the degradation constant in peak fitting algorithm. A numerical study for the analysis of the Doppler-broadened line shape has been conducted by several groups. Kubo and Sakai [3] present a simplified equation without consideration of the detector response as below

$$g(E) = q \left[1 - \left(\frac{|E - E_0|}{E_0 v_0 / c} \right)^{\frac{\lambda - D}{D}} \right], \quad |\Delta E| \leq \Delta E_m \quad (1)$$

where E is the shifted energy by Doppler-broadening, c is the speed of light, E_0 is the emitted gamma-ray energy, ΔE is the Doppler shift given to be 7.6 keV for ^7Li ion, v_0 is the initial velocity of recoiling ^7Li ion, q is the height of the spectrum at $E = E_0$ and is related to the number of emitted gamma-rays N_0 by $q = \frac{cN_0}{2E_0v_0} \frac{\lambda}{\lambda - D}$, λ is the decay constant of $^7\text{Li}^*$ and D is a degradation constant which depends on the stopping power of the slowing down medium.

Further approach has been done by Choi *et al.* [7], which includes a routine performing parametric search and fitting the measured spectrum in a least squares method. This includes the convolution integral between the emission spectrum and the detector response function based on the peak fitting routine in a HYPERMET [8] and was implemented in a recent version of HYPERGAM code [9].

2. Experimental

2.1 Sample preparation

Boron samples used in this study were made from boric acid powder, elemental boron, boron carbide, titanium di-

* Author for correspondence (E-mail: gmsun@kaeri.re.kr).

boride, La-hexaboride, Pr-hexaboride, Nd-hexaboride, Li-pentaborate, and industrial boron steel chips. The sample was sealed into a cylindrical Teflon vial and mounted into the Teflon holder located on the center line of the neutron beam and then irradiated.

2.2 Prompt gamma-ray measurement

Prompt gamma-ray spectra were measured at the PGAA facility at the HANARO of the Korea Atomic Energy Research Institute [10–12]. The true integrated neutron flux at the sample position was 1.0×10^8 n/cm² s with a 30 MW thermal power operation and the cadmium ratio for Au was 266. The effective neutron temperature is 269 K at the sample position. The gamma-ray spectrometer system consists of an n-type HPGe detector with a relative efficiency of 43% and a shield detector of a BGO/NaI(Tl) scintillator placed 25 cm away from the sample with respect to the neutron beam direction. The prompt gamma-ray spectra were accumulated in the Compton suppression mode [13]. A detailed off-line analysis was carried out by the HYPGAM including a Doppler-broadened boron peak (DBP) analysis routine.

2.3 Implementation of DBP algorithm into HYPGAM code

The Choi's algorithm was implemented into HYPGAM coded by MATLAB and its graphical user interface is shown

in Fig. 1. It includes many parameters for DBP line shape with tail and step terms, continuum background and several interfering Gaussian gamma-ray peaks originating from impurities or other component elements in the samples. More details about the code were already dealt with in another article [7]. In the figure, the α parameter is related to the degradation constant as in Eq. (2).

$$\alpha \equiv \frac{\lambda}{D} - 1 \quad (2)$$

3. Results and discussion

The spectrum of Doppler-broadened gamma rays emitted from ⁷Li* during slowing down in a stopping medium is decomposed as shown in Fig. 2, where nine spectra were shown for elemental boron (a), boric acid (b) and boron carbide (c), titanium diboride (d), La-hexaboride (e) and Pr-hexaboride (f), Nd-hexaboride (g), Li-pentaborate (h) and boron steel (i). The fitting region is fixed between 460 keV ~ 495 keV, which was linearly calibrated from the channel-energy calibration process using a multiline gamma-ray standard source. The Doppler shift is fixed as 7.6 keV and the full width at a half maximum for the 472.2 keV interfering line emitted from the ²⁴Na produced by a neutron capture of ²³Na is given as around 1.3 keV for all spectra. The algorithm can search for all the parameters at once, but

Fig. 1. User interface of the DBP line analysis. Many parameters for DBP line with tail and step terms, continuum background and interfering Gaussian peaks.

Fig. 2. Analysis of DBP line for boric acid sample for elemental boron (a), boric acid (b) and boron carbide (c), titanium boride (d), La-hexaboride (e) and Pr-hexaboride (f), Nd-hexaboride (g), Li-pentaborate (h) and boron steel (i). For simplicity, only total background line, step background line, DBP line and fitted line are shown with the raw data.

since there are so many parameters over ten, we used batch run, which consists of several fitting processes searching for several parameters per each fitting. At first, we searched for less sensitive parameters for peak center position, amplitude and widths of the peaks, the continuum background, and then for more sensitive ones for tail and step background to each peak and for α factor seriously depending on the stopping media of Li ion. This skill is very successful. By fitting the spectrum, it is clearly distinguishable as shown in the figures that the line shapes of the DBP and its interfered gamma-ray peaks are decomposed by using the new algorithm. The relative portion of the count rate under the short tails to the DBP line and Gaussian lines for the total peak count rate was very small and not seen here.

The α factors and degradation constants (D) for various media were obtained as a result of multi-parametric fitting process and listed in Table 1. The uncertainties of all the parameters including α factor and degradation constant are calculated through a reduced chi-square and a Hessian matrix based on gradient of fitting function and numerical differentiation. [14]

The degradation constant of Li ions reflects the stopping power of the environmental matrix around the boron. The constant value for boric acid with a density of 1.435 g/cm^3

Table 1. α factor and degradation constant(D) for various media.

Samples	α , ps	Degradation constant D , ps^{-1}
Boric acid	6.904	1.201 ± 0.028
Elemental boron	6.376	1.287 ± 0.024
Boron carbide	3.419	2.147 ± 0.016
Titanium diboride	3.670	2.032 ± 0.022
La-hexaboride	3.671	2.032 ± 0.020
Pr-hexaboride	3.495	2.110 ± 0.019
Nd-hexaboride	3.496	2.111 ± 0.020
Li-pentaborate	3.495	2.111 ± 0.019
Boron steel	2.744	2.535 ± 0.016

is determined to be $1.201 \pm 0.028 \text{ ps}^{-1}$ in this study and the theoretical value by using SRIM is 1.34 ps^{-1} [7]. In other works by Sakai *et al.* [2] and Yamauchi *et al.* [15], degradation constant was determined to be 1.22 ps^{-1} for aqueous boric acid solution and 1.48 ps^{-1} for polycrystalline boric acid, respectively.

Fig. 3 shows normalized peak lines for three samples with different degradation constant. Elemental boron, titanium diboride and boron steel were selected for distinct comparison of the line shape. The sample with the higher degradation constant shows the sharper line shape.

Fig. 3. Normalize peak lines for various media with different degradation constant.

4. Conclusion

In this study, the Doppler-broadened line shapes from the 9 boron-containing samples with different matrix and different stopping power were analyzed by using a new Choi's algorithm which deals with the line shape as a convolution integral of the detector response function and the emission gamma-ray spectrum from the recoiled ion. The degradation phenomenon of the Li ion in the stopping media is inserted as fitting parameter. Rare-earth borides show similar degradation constant larger than another samples. For the media with low stopping power like elemental boron, boric acid powder, the degradation constants are less than 1.3 ps^{-1} . However Li-pentaborate with similar mass with elemental boron shows similar value with rare-earth borides, which has to be further studied in detail. And for the gamma rays emitted from boron atom in the steel sample, the D value is 2.535 ps^{-1} .

Acknowledgment. This research was supported by the Nuclear R&D Program by the Ministry of Science and Technology, Republic of Korea.

References

1. Hauser, U., Neurwirth, W., Pietsch, W., Richter, K.: On the determination of collision cross sections by nuclear Doppler shift. *Z. Phys. A* **269**, 181 (1974).
2. Sakai, Y., Yonezawa, C., Matsue, M., Magara, M., Sawahata, H., Ito, Y.: Measurement and analysis of the line shape of prompt γ -rays from recoiling ^7Li produced in the $^{10}\text{B}(n, \alpha)^7\text{Li}$ reaction. *Nucl. Instrum. Methods A* **353**, 699 (1994).
3. Kubo, M. K., Sakai, Y.: A simple derivation of the formula of the Doppler broadened 478 keV γ -ray lineshape from ^7Li and its analytical application. *J. Nucl. Radiochem. Sci.* **1**, 83 (2000).
4. Magara, M., Yonezawa, C.: Decomposition of prompt gamma-ray spectra including the Doppler-broadened peak for boron determination. *Nucl. Instrum. Methods A* **411**, 130 (1998).
5. Sun, G. M., Chung, Y. S., Park, C. S., Choi, H. D.: Doppler-broadened boron peak analysis by using a modified spectral decomposition algorithm. *J. Radioanal. Nucl. Chem.* **278**, 637 (2008).
6. Firestone, R. B., Shirley, V. S., Baglin, C. M., Frank, S. Y., Zipkin, J.: *Table of Isotopes*. 8th Edn., John Wiley & Sons Inc., New York (1996).
7. Choi, H. D., Jung, N. S., Park, B. G.: Analysis of Doppler-broadened peak in thermal neutron induced $^{10}\text{B}(n, \alpha\gamma)^7\text{Li}$ reaction by using HyperGam. *Nucl. Eng. Techn.* **41**, 113 (2009).
8. Phillips, G. W., Marlow, K. W.: Program HYPERMET for Automatic Analysis of Gamma-ray Spectra from Germanium Detectors. NRL Memorandum Report 3198 (January 1976).
9. Park, C. S., Choi, H. D., Sun, G. M., Whang, J. H.: Status of developing HPGe γ -ray spectrum analysis code HYPERGAM. *Prog. Nucl. Energy* **50**, 389 (2008).
10. Byun, S. H., Sun, G. M., Choi, H. D.: Beam characteristics of polychromatic diffracted neutrons used for prompt gamma activation analysis. *J. Korean Nucl. Soc.* **34**, 30 (2002).
11. Byun, S. H., Sun, G. M., Choi, H. D.: Development of a prompt gamma activation analysis facility using diffracted polychromatic neutron beam. *Nucl. Instrum. Methods A* **487**, 521 (2002).
12. Byun, S. H., Sun, G. M., Choi, H. D.: Characterization of a polychromatic neutron beam diffracted by pyrolytic graphite crystals. *Nucl. Instrum. Methods A* **490**, 538 (2002).
13. Sun, G. M., Park, C. S., Choi, H. D.: Performance of a Compton Suppression Spectrometer of the SNU-KAERI PGAA facility. *J. Korean Nucl. Soc.* **35**, 347 (2003).
14. Press, W. H., Teukolsky, S. A., Vetterling, W. T., Flannery, B. P.: *Numerical Recipes*. Cambridge University Press (1989), p. 523.
15. Yamauchi, S., Sakai, Y., Watanabe, Y., Kubo, M. K., Matsue, H.: Analysis of boron in wood treated with boric acid solutions using Doppler broadening method of prompt gamma-rays. *J. Wood Sci.* **1** (2006).